

Pledge and signature:

Note: If you want your paper returned folded (*i.e.*, score concealed), please print your name on the back.

1. (3) In an adsorption experiment, the manifold volume is calibrated with a bulb having $V = 34.5 \text{ cm}^3$. Gas having $P = 207 \text{ Torr}$ is expanded from the bulb to the rest of the system (previously evacuated), giving $P = 79.3 \text{ Torr}$. Calculate the volume of the rest of the system.

2. (3) The volume just calculated includes a small connecting region between the calibration bulb and the manifold. Gas of $P = 195.2 \text{ Torr}$ is trapped in this section and then expanded into the manifold (previously evacuated), giving $P = 19.2 \text{ Torr}$. Calculate the volumes of (a) the connecting region and (b) the manifold.

3. (4) In a different system, a sample cell of volume 13.3 cm^3 is connected to a manifold having $V = 63.5 \text{ cm}^3$. The system is initially at $T = 299 \text{ K}$ with N_2 gas at $P = 178.3 \text{ Torr}$. Then the flat part of the sample cell is immersed in liquid N_2 (77.0 K) and the pressure drops to 139.2 Torr . Calculate the "cold volume."

4. (6) Adsorption data for the adsorbed amount v (STP cm^3) can be analyzed by fitting to two different relationships, one of which gives a straight-line presentation.
 - (a) Give the quantities "y" to be taken as dependent variable in each of these fits.
 - (b) Assuming the measured v s have constant uncertainty, how should the data in each of these two fits be weighted?
 - (c) If these data have proportional uncertainty ($\propto v$), how should the data be weighted in each case?