



Figure 5. VP-ITC temperature correction (true - fiducial) over the stated range. The curve is a fitted quadratic in the argument $(t - 25^{\circ})$ and gives an error of 0.123(3) K at 25°C. Most values beyond 25° were recorded after the VP-ITC instrument had equilibrated at the specified temperature.